

Technical Evaluation Report for the Purchase of Category-A Equipment																															
Sl	Firm Name	Name of equipment	Product Evaluation Parameter													Firm Evaluation Parameters										Technical Marks					
			Conformance To Specification			Product International Certification			Manufacturer Performance			After Sale Product Local Performance			Warranty	Personnel/ Human Resource	Workshop facility			Firm Financial Strength			Firm Registration		Technical Marks (70)	Remarks					
			Fully compliance with the required specifications as per Statement of Requirements. If not, the condition that marks functions and performance in any aspect would not be affected. More than 4 minor deviations will be considered as major deviation and the bidder will be considered as non-responsive for the quoted item. (One mark for each deviation with the sub-points)	Certificate of US Food and Drug Administration (FDA) of the quoted model.	Certificate of European community medical device directives (CE mark) for the quoted model.	AS/NZS 60601 for the quoted model. Issued by Verified Certification Body.	Certification of Conformity, CE, Technical Certificate of Ministry of health and welfare Japan (MHWR) of the quoted model. (Translated English Version)	Valid ISO 13485 Medical Device Quality Management System Certificate.	ORHAS 8001 / ISO 4001 Occupational Health & Safety Series from IAP accredited body.	Weightage for local Pakistan manufacturer.	One mark for each major manufacturer performance certificate (validity of the firm on letter head signed and stamped later for the quoted model) or product of Pakistan. Performance certificate shall be copied with supply order of purchase order from public sector medical institutions. Supply Order / Purchase Order / Installation certificate shall be considered as evidence of performance certificate.	One mark for each after sale customer performance certificate (validity of the equipment from the existing level research level private sector medical institutions of Pakistan. The hospital must be recognized from Pakistan Medical Council (PMC) for the quoted model. Performance certificate shall be copied with supply order / purchase order from public sector medical institutions. Supply Order / Purchase Order / Installation certificate shall be considered as evidence of performance certificate.	Minimum Period of these parts, both with spare parts and services. Note: this year services only without parts, from the date of installation. (Combining average of electrical/electronic, mechanical, pneumatic, hydraulic, and other parts).	Diploma of Associate Engineer (DAE) in electrical / electronic / biomedical / mechanical / industrial. PEC registration card of the engineer must be submitted. (One mark for each certificate)	Graduate Engineer with PEC Registration in electrical / electronics, biomedical / mechanical / industrial. PEC registration card of the engineer must be submitted. (One mark for each certificate)	Availability of workshop in higher Pakistans to be verified with Owners / Client Agreement with Owner / Rent Agreement with Company name.	List of colored tools available at workshop / List of Printing and Calibration tools for the quoted item available at workshop / Detail list of Spare parts available at workshop for the quoted item. (1 mark each)	Annual Sales tax returns for last two years (1 mark for each year).	Annual Income tax returns for last two years (1 mark for each year).	Last two years Audited Balance Sheet / Chartered Accountant / Auditor's report for last 2 years. (1 mark for each year).	PEC registration certificate of the firm from PNAC.	Firm registered with DPA/PEC (Drug Regulatory Authority of Pakistan) to import / manufacture of medical devices.	Valid ISO 9001 Quality Management Certificate of the firm from PNAC.	Total Technical Score							
			30	0	0	0	3	2	0	0	0	4	3	Mandatory	2	4	Mandatory	2	2	2	2	2	2	2	2	2	2	2	70		
1	M/S Mercy Enterprises	Autoclave 50L Electric Model: BKQ B501 Make: BioBase China	0	0	Yes	0	3	0	0	0	0	4	0	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	19	Non Responsive 1.High vacuum function not available in the quoted model		
1	M/S Matara Engineering Corporation	Autoclave 50L Electric Model: HS-VSP 50 Make: Hongshun Medical China	0	0	Yes	0	3	0	0	0	0	0	0	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	19	Non Responsive 1.High vacuum function not available in the quoted model		
1	M/S Shah's brothers and Co	Autoclave 50L Electric Model: autoclave vertical 50 liters Make: jiangsu shemong Autoclave inc China	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive 1. One out of three mandatory certificates not available.
1	M/S ideal business products	Autoclave 50L Electric Model: LS-50HW Make: Jiangsu binjiang China	0	0	Yes	0	3	0	0	0	2	2	1	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	22	Non Responsive 1.High vacuum function not available in the quoted model		
1	M/S Friends Traders	Autoclave 50L Electric Model: Most-T Order Code T45 Make: Shiva China	30	0	Yes	0	3	2	0	0	2	2	0	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	53	Responsive		
1	M/S MKJ associates	Autoclave 50L Electric Model: DW-B 50 L Make: Shanghai Drexell scientific instruments China	0	0	Yes	0	0	0	0	0	0	0	0	0	Yes	2	4	Yes	2	2	2	2	2	2	Yes	Yes	2	16	Non responsive 1. Not according to the Approved specification manual quoted instead of automatic		
1	M/S Alam Medix	Autoclave 50L Electric Model: LX-B Make: MT Medical China	0	0	Yes	0	3	0	0	0	2	2	1	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	22	Non Responsive 1.High vacuum function not available in the quoted model		
2	M/S Medco	Baby Phototherapy Model: BL-500 Make: Zhengzhou diosen Instrument	30	0	Yes	0	3	2	0	0	0	0	0	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	51	Responsive		
2	M/S Medbridge	Baby Phototherapy Model: BBP-2000B Make: Suzhou Being China	30	0	Yes	0	3	0	0	0	0	0	0	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	49	Responsive		
2	M/S ideal business products	Baby Phototherapy Model: B-100(LED) Make: Shanghai Ro Chain China	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive 1. Light source not as per tender requirements	
2	M/S Friends Traders	Baby phototherapy Model: BL-70 Make: Comen Medical China	30	0	Yes	0	3	2	0	0	0	0	0	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	51	Responsive		
2	M/S MKJ associates	Baby Phototherapy Model: BT 400 Make: Bistos Korea	30	Yes	3	0	3	2	0	4	3	3	3	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	61	Responsive		
2	M/S Alam Medix	Baby Phototherapy Model: BN300 Make: BNG Medical China	30	0	Yes	0	3	2	0	4	3	3	3	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	58	Responsive		
2	M/S Medifa Enterprises	Baby Phototherapy Model: BBP-2000B Make: Suzhou Being China	30	0	Yes	0	3	0	0	1	0	0	0	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	50	Responsive		
2	M/S Paradise Export Company	Baby Phototherapy Model: BT-400 Make: Bistos Korea	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non responsive 1. Mandatory Documents i.e PEC registration Deap not available	
3	M/S Medco	Baby Suction Machine Model: DFX-23D1 Make: Jiangsu Kelong Medical	30	0	Yes	0	3	2	0	0	0	0	0	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	51	Responsive		
3	M/S Suda Associates	Baby Suction Machine Model: MS-700 Make: M Technology Korea	30	0	Yes	0	3	0	0	4	3	3	3	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	56	Responsive		
3	M/S Shah's brothers and Co	Baby Suction Machine Model: Baby suction apparatus Make: Yuwell China	30	0	Yes	0	3	2	0	0	0	0	1	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	52	Non Responsive 1.20L with jar quoted instead of 30L		
3	M/S ideal business products	Baby Suction Machine Model: SS-8A Make: Shanghai medical China	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive 1. Quoted 20 L, instead of 30L	
3	M/S Friends Traders	Baby Suction Machine Model: DFX-500K Make: Doctor's Friends China	30	0	Yes	0	3	0	0	0	0	0	0	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	49	Responsive		
3	M/S MKJ associates	Baby Suction machine Model: 7a-23b Make: Jiangsu Yuwell Medical China	30	0	Yes	0	3	2	0	4	3	3	3	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	58	Responsive		
3	M/S Alam Medix	Baby Suction Machine Model: DFX-23D Make: MT Medical china	30	0	Yes	0	3	0	0	4	3	3	3	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	56	Responsive		
3	M/S Medifa Enterprises	Baby Suction Machine Model: BES-9030C Make: Suzhou Being Medical China	30	0	Yes	0	3	0	0	1	0	0	0	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	50	Responsive		
3	M/S Paradise Export Company	Baby Suction Machine Model: 6.4 P Make: 3A Healthcare	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non responsive 1. Mandatory Documents i.e PEC registration Deap not available	
3	M/S Ansha international	Baby Suction Machine Model: SU-305P Make: Genny Taiwan	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive 1. Mandatory certificate Deap/Pakistan engineering council not available	
4	M/S Medline Technology	CTG machine Model: T10 Make: Huarime China	30	0	Yes	0	3	0	0	4	0	4	0	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	53	Responsive		
4	M/S Medbridge	CTG machine Model: MCF-21K Make: Shenzhen Unicare China	30	0	Yes	0	3	0	0	0	0	0	0	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	49	Responsive		
4	M/S Suda Associates	CTG machine Model: MCF-21K Make: Shenzhen Unicare China	30	0	Yes	0	3	2	0	4	3	3	3	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	58	Responsive		
4	M/S Friends Traders	CTG Machine Model: C21 Make: Comen China	30	0	Yes	0	3	2	0	0	0	0	0	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	51	Responsive		
4	M/S MKJ associates	CTG Machine Model: BT-150 Make: Bistos Korea	30	Yes	3	0	3	2	0	4	3	3	3	Yes	2	4	Yes	2	2	2	2	2	2	2	Yes	Yes	2	61	Responsive		

