

Directorate General health Services Khyber Pakhtunkhwa
 Health Department, Khyber Pakhtunkhwa, Peshawar
 Draft Bid Evaluation report (Technical score) of Equipment Tender for COVID-19 Emergency in 4th Wave of COVID-19.

All concerned to please note in the public interest that;

1. This list is a draft tentative list to serve the only purpose of ensuring more transparency and equity in the bidding process by providing equal opportunities to all concerned bidders to have access to information related to their bids for Healthy & fair competition; and
2. This list will be kept on this website till 5:00 p.m. on 5th of August 2021 (Thursday); and
3. This list is not an approved legal document to have any authority or weightage or value for any legal proceedings in any court of law or any other forum; and
4. This list is being shared to eliminate the possibility of any error in terms of figures or words related to Technical Bids; and
5. It is stressed to please note that this list is in no way a final document to be used as an evidence or argument for selection of products for rate contracting from the lowest bidder or award or justification of supply contract / order to the selected bidder by the procuring agency; and
6. All possible efforts have been made to reduce human errors and omissions but still inadvertent errors and omissions may possibly be expected in this list. All the bidders participating in this bidding competition are expected and requested to thoroughly examine this and point out any error or omission and total marks obtained, etc.
7. Any feedback related to errors and omissions must reach the office of DGHS, Ex fata Secretariat Warsak Road Peshawar not later than 5.00 p.m. on 5th of August 2021 (Thursday), in written form duly signed and attested by the responsible relevant person of the bidding entity and
8. This list is only meant to rectify / correct the mistakes, if any that might have occurred during the data entry process for the already submitted Technical Bids to facilitate final bidding competition by opening Financial Bids of Responsive bidders to be held shortly by the High Power Purchase Committee for COVID-19 Emergency; and
9. Any comments, feedback etc. received after the above mentioned cutoff time and date (Para-7 above) will not be considered and the so finalized list will be placed before the High Power Purchase Committee for COVID-19 Emergency Procurement in the best public interest.
10. The High Powered Purchase committee, at its discretion as per KPPRA rules may decide select or otherwise any item mentioned in this BER for the current bidding competition in Emergency Procurement of COVID-19 4th wave.
11. This Technical bids of bidder who participated in bidding competition are evaluated in accordance with already advertised evaluation criteria, specification and terms & Condition.

S.No	Firm Name	Equipment Name Model, Make	Compliance with Specs	Mandatory Certifications	Warranty	Authorization	Available Quantity	Delivery Time	Remarks
------	-----------	----------------------------	-----------------------	--------------------------	----------	---------------	--------------------	---------------	---------

1	Hospicare Systems	Syringe Pump Model: Perfuser Compact Plus Make:B.Braun Germany	Yes	CEMDD,ISO 13485	3+2 years	Yes	0	2 in 8-10 weeks	Responsive
2	Hospicare Systems	ECG 3 Channel Model: 100L Make:Cardioline italy	Yes	CEMDD,ISO 13485	3+2 Years	Yes	0	4 weeks after confirm order	Responsive
3	Hospicare Systems	Centrifuge Model:Sorvall st-8 Make:Thermofisher scientiufic USA	Yes	CEMDD	3+2 Years	Yes	2	Ready Stock	Responsive
4	Hospicare Systems	HFNC Model:HFNO Make:Alson Pakistan	Yes	No	3+2 Years	Yes	250	50 units by Mid of august 50 units by End of august	Non Responsive 1.No CEMDD ,USFDA Available
5	Hospicare Systems	Cardiac Monitor Model:Q3 Make:Biolight China	Yes	CEMDD,US FDA,ISO 13485	3+2 Years	Yes	250-300	3 to 4 weeks	Responsive
6	Hospicare Systems	Ventilator Model:PB840 Make:Medtronic USA Country of Origin:Ireland	Yes	USFDA,CE MDD,ISO 13485	3+2 Years	Yes	8	8 in 2-3 days 40 units within 6-8 weeks,120 units by Feb 2022	Responsive

7	Hospicare Systems	CPAP Model:HFNO Make:Alson Pakistan	Yes	No	3+2 Years	Yes	250	50 units by Mid of august 50 units by End of august	Non Responsive 1.No CEMDD ,USFDA Available
8	Hospicare Systems	ABG Machine Test 15000 and 10 ABG Model:ISTAT along with CG4 Cartridge Make: ABBOTT USA	Yes	No CEMDD,US FDA ISO 13485 availble	3+2 Years	Yes	10 ABG along with 15000 TEST	In Stock	Non Responsive 1.USFDA not Available. 2.CEMDD not available
9	Hospicare Systems	ABG Machine Model:I-STAT Make: ABBOTT USA	Yes	No CEMDD,US FDA ISO 13485 availble	3+2 Years	Yes	6	Ready in Stock. Remaining in 4-6 weeks	Non Responsive 1.USFDA not Available. 2.CEMDD not available
10	Med Express	Ventilator Model: Evolution 3e-T Make: Event USA	Yes	USFDA,CE MDD	3+2 Years	Yes	100	5-6 weeks	Responsive
11	Med Express	HFNC Model: NF-5 Make:Comen China	Yes	CEMDD,ISO 14001,ISO 45001,ISO 9001	3+2 Years	Yes	Null	4-5 weeks	Responsive
12	Med Express	Cardiac Monitor Model:C-80 Make:Comen China	Yes	USFDA,CE MDD,ISO 9001,ISO14001,ISO45001	3+2 Years	Yes	Null	4-5 weeks	Responsive

13	Med Express	Syringe Pump Model: M200A Make:Comen China	Yes	CEMDD,ISO 9001,ISO14 001,ISO450 01	3+2 Years	Yes	26	26 in 1 day remaining in 4- 5 weeks	Responsive
14	Friends Traders	Ventilator Model:SV 300 Make:Mindray china	Yes	CEMDD,M HLW,ISO13 485,Iso 9001	3+2 Years	Yes	6	6 units in 1 day remaining in 45 days	Responsive
15	Friends Traders	Defibrillator Model:D3 Make:Mindray China	Yes	CEMDD,M HLW,ISO13 485,Iso 9001	3+2 Years	Yes	15	15 units in 10 days Remaining upto 24th august	Responsive
16	Friends Traders	Cardiac MonitorModel:Umec 12 Make:Mindray China	Yes	CEMDD,US FDA,MHL W,ISO`s	3+2 Years	Yes	20	20 units to 10 august remaing in 60 working days	Responsive
17	MKJ Associates	Cardiac Monitor Model: CMS8000 Make:Contec China	Yes	CEMDD,US FDA, ISO`s	3+2 Years	Yes	500	100 in 12 days rest in 15 to 25 days	Responsive
18	MKJ Associates	ECG 3 Channel Model: ECG300GA Make:Contec Medical China	Yes	CEMDD,US FDA, ISO`s	3+2 Years	Yes	25	Ready stock	Responsive
19	MKJ Associates	HFNC Model: Sleepas Elite Make:Foras Medical Turkey	Yes	CEMDD,IS O 13485	3+2 Years	Yes	120	120 Units 15-25 days	Responsive

20	MKJ Associates	Bipap Model: Yel Bi PAP S/T Make:Foras Medical Turkey	Yes	CEMDD,ISO 13486	3+2 Years	Yes	300	15-25 days	Responsive
21	Matora Engineering Corporation	Ultrasound Machine Model:Cbit 8 Make:Chison China	Yes	CEMDD,US FDA	3+2 Years	Yes	Nil	25 Days	Responsive
22	Matora Engineering Corporation	Oxygen tank 5000 L Model:CFL 5/1.6/A Make:Shangdong Zhingjie Make:China	No	-	3+2 years	Yes	Nil	90 Days	Non Responsive 1.Capacity Less than Required
23	Matora Engineering Corporation	HFNC Model:HFO-1 Make:Zhejiang LifeMed China	Yes	CEMDD,ISO 13485	3+2 Years	Yes	Nil	20 Days	Responsive
24	Matora Engineering Corporation	Portable X ray Machine Model;SM-32-HF-Batt Make:sedecal Spain	Yes	CEMDD,ISO	3+2 Years	Yes	25	90 days after supply order	Non Responsive 1.USFDA attached is not for the quoted Model
25	Matora Engineering Corporation	Oxygen tank 10000 L Model:CFL 10/1.6/A Make:Shangdong Zhingjie China	No	-	3+2 years	Yes	Nil	90 Days	Non Responsive 1.Capacity Less than Required

26	FTS	Bipap Model:25s/25A/25ST/25STA Make:resvent China	Yes	CEMDD,ISO	3+2 Years	Yes	33	3 Days Remaining in 7-10 days	Responsive
27	FTS	Cpap Model:20C/20C Pro Make:resvent China	Yes	CEMDD,ISO	3+2 years	Yes	75	4 Days Remaining in 7-10 days	Responsive
28	Hoor Pharma	ABG Machine Model:Rapid Lab 348Ex Make:seimens UK	Yes	ISO	3+2 years	Yes	10	2 Days	Non Responsive 1.Certificate of Confirmity attached instead of CEMDD. 2.The attached FDA is enlismnt not USFDA
29	Hoor Pharma	ABG Machine with 15000 test and 10 Machines Model:Rapid Point 500 Make:seimens USA/EU	Yes	ISO	3+2 years	Yes	Nil	2-3 Weeks	Non Responsive 1.Certificate of Confirmity attached instead of CEMDD. 2.The attached FDA is enlismnt not USFDA

30	Hoor Pharma	Portable X ray Machine Model:Mobile Art Evolution Make:Shimadzu Japan	Yes	CEMDD,M HLW,USFD A	3+2 years	Yes	2	2 units in 2 days and rest in 4 to 8 weeks	Responsive
31	Global Healthcare	ABG Machine Model:Opti CCA TS2 Make:Opti Medicsal USA	Yes	Nil	3+2 years	Yes	Nil	6-8 weeks after confirm order	Non Responsive 1.Authorizat ion Certificate is awarded to Global Marketing not to Global Healthcare . 2.USFDA not available. 3.Certificate of confirmity attached instead of CEMDD
32	Global Healthcare	ABG Machine with 15000 test and 10 Machines Model:Opti CCA TS2 Make:Opti Medicsal USA	Yes	Nil	3+2 years	Yes	Nil	6-8 weeks after confirm order	Non Responsive 1.Authorizat ion Certificate is awarded to Global Marketing not to Global Healthcare . 2.USFDA

									not available. 3.Certificate of confirmity attached instead of CEMDD
33	Global Healthcare	Syringe Pump Model: Hy-900B Make:guanzhou Beaconn Medical China	Yes	CEMDD	3+2 years	Yes	15	6-8 weeks after confirm order	Responsive
34	Radiant Medical Pvt Ltd	Syringe Pump Model:CS-20 Make:Care Prologic Medical China	Yes	CE,ISO 13485	3+2 years	Yes	62	1-2 weeks	Responsive
35	HealthCare Services	Cardiac Monitor Model:CPR 9000-A Make: China	Yes	ISO	3+2 years	Yes	200	2-3 Hours remaining in 30 days	Non Responsive. 1.USFDA not available.
36	Bio Tech Services	Defibrilator Model:TEC-5631 Make:Nihon Kohden Japan	Yes	CEMDD,M HLW	3+2 years	Yes	14	2-3 working days	Responsive
37	Bio Tech Services	Cardiac Monitor Model:BSM 3562 Make:Nihon Kohden Japan	Yes	CEMDD,M HLW	3+2 years	Yes	52	2-3 days	Responsive

38	Bio Tech Services	ECG 3 Channel Model:ECG 305 Make: Suzuken japan	Yes	MHLW,CE	3+2 years	Yes	21	2-3 days	Responsive
39	Bio Tech Services	Ultrasound Machine Model: HS-2600 Make:Honda Electronics Japan	No	MHLW,CE	3+2 years	Yes	17	2-3 days	Non Responsive 1.The qouted model Has 2 probes instead of 3.
40	Mercy Enterprises	HFNC Model:Humid-BM Make:Respicare China	Yes	CE	3+2 years	Yes	15	1 Hour Remaining in 20 days	Responsive
41	Agfa Healthcare	Portable X Ray Model:Topaz 32D Make:Drgem Korea	Yes	CE,USFDA	3+2 years	Yes	25	40 Days	Responsive
42	Ideal Bussiness Products	Ventilator model:V8800 Make:Oricare China	Yes	CEMDD,US FDA,ISO 13485	3+2 years	Yes	186	In November	Responsive
43	Ideal Bussiness Products	Cardiac Monitor Model:G3D Make:General Meditec China	Yes	CEMDD,US FDA	3+2 years	Yes	Null	200 in 25 days	Responsive
44	Ideal Bussiness Products	Oxygen Tank 10000 Model:CDL-10/1.6 Make:zhuhai senbo China	Yes		3+2 years	Yes	Null	20 unit 90-120 days	Responsive

45	Ideal Bussiness Products	Oxygen Tank 5000 Model:CDL-5/1.6 Make:zhuhai senbo China	Yes		3+2 years	Yes	Nil	20 unit 90-120 days	Responsive
46	Ideal Bussiness Products	Ultrasound Machine Model:Qbit 5 Make:Chison Medical china	No	USFDA,CE MDD,ISO	3+2 years	Yes	Nil	10 in 20 days	Non Responsive 1.the Qouted model Has 2 active probes instead of 3.
47	Ideal Bussiness Products	Syringe Pump Model:MS-31 Mdkingdom Make:Zhejiang China	Yes	CEMDD,IS O 13485	3+2 years	Yes	Nil	100 in 20 days	Responsive
48	Ideal Bussiness Products	HFNC Model:HFO-1 Make:Zhejiang LifeMed China	Yes	CEMDD,IS O	3+2 years	Yes	Nil	250 in 20 days	Responsive
49	Ideal Bussiness Products	Bipap Model:BF30ST Make:BEIJING AeonMed China	Yes	CEMDD,IS O	3+2 years	Yes	Nil	160 in 15 days	Responsive
50	Ideal Bussiness Products	Bipap Model:La25B Make: Zhejiang Lifemed China	Yes	CeMDD,IS O	3+2 years	Yes	Nil	200 in 15 Days	Responsive
51	Endoaid BioMedica	ECG 3 Channel Model:EKG I Make: I-G Medical Germany	Yes	CEMDD,IS O	3+2 years	Yes	Nil	45-60 Days	Responsive

52	Shirazi Trading	Ventilator Model:R860 Make:Gehealthcare China	Yes	USFDA,CE MDD	3+2 years	Yes	Null	50 in 8-10 weeks remaining 12- 16 weeks	Responsive
53	Shirazi Trading	Cardiac Monitor Model: B105 Make:GE healthcare China	Yes	USFDA,CE MDD	3+2 years	Yes	Null	50 in 8-10 weeks remaining 12- 16 weeks	Responsive
54	Shirazi Trading	Portable X ray Model: Traveller Make: GE Healthcare Italy	Yes	CEMDD	3+2 years	Yes	Null	3-4 Months	Non Responsive 1.Two mandatory Certificates not available
55	FDS PVT LTD	Ventilator Model:R30 Plus Make:Siriusmed Medical China	Yes	CEMDD,IS O 13485	3+2 years	Yes	Null	186 in 8-10 weeks	Non Responsive 1.Two Mandatory certificates not available
56	FDS PVT LTD	Ventilator Model:R50 Plus Make:Siriusmed Medical China	Yes	CEMDD,IS O 13486	3+2 years	Yes	Null	186 in 8-10 weeks	Non Responsive 1.Two Mandatory certificates not available
57	FDS PVT LTD	Bipap Model: DM28-25ST-BP Make: Yamid Medical China	Yes	CEMDD,IS O 13485	3+2 years	Yes	Null	500 in 08-10 weeks	Responsive

58	Tahir Scientific	ABG Machine Model:Nova stat ,pHOX Co Oximetry Make: Nova Biomedical USA	Yes	No CEMDD, No USFDA	3+2 years	Yes	2	2 ready stock remaining in 45 Days	Non Responsive 1.No Mandatory Certificates Available. 2.the Qouted Model is Nova stat profile Co-Oximetry while the USFDA attached is of the model Nova stat pHOX plus
59	Tahir Scientific	ABG Machine with 15000 test and 10 machines Model:Nova stat ,pHOX Co Oximetry Make: Nova Biomedical USA	Yes	No CEMDD, No USFDA	3+2 years	Yes	2	3 ready stock remaining in 45 Days	Non Responsive 1.No Mandatory Certificates Available. 2.the Qouted Model is Nova stat profile Co-Oximetry while the USFDA attached is of the model Nova stat pHOX plus
60	Shahco Medical PVt	Ultrasound Machine Model:S-11 Plus Make:Sonoscape Medical China	Yes	CEMDD,ISO	3+2 years	Yes	5	5 in days,remaining in 4 weeks	Non Responsive 1.Two mandatory Certificates not available

61	Ghani Chemical Industries	Oxygen Tank 10000 with local Vaporizer Model:not Mentioned Make:Not Mentioned	Yes		3+2 years	Yes	2	Ready Stock	Responsive
62	Ghani Chemical Industries	Oxygen Tank 5000 with local Vaporizer Model:not Mentioned Make:Not Mentioned	Yes		3+2 years	Yes	4	Ready Stock	Responsive
63	Ijaz enterprises	Ventilators Model:S11000 Make:Superstar China	Yes	No USFDA,MH LW available	3+2 years	Yes	186	50 units in 10-15 days ,136 in 25-35 days	Non Responsive 1.Two out of Three Mandatory Certificates not Available.
64	Ijaz enterprises	Cardiac Monitor Model:Spr900B Make:DOTT Medical China	Yes	No USFDA,MH LW available	3+2 years	Yes	906	300 in 7 Days remaining in 25-35 days	Non Responsive 1.Two out of Three Mandatory Certificates not Available.
65	Ijaz enterprises	HFNC Model: Not mentioned Make: Guangdong pigon China	Yes	No CeMDD, No USFDA, No MHLW	3+2 years	Yes	250	100 Units in 7-14 Days remaining in 15 to 25 days	Non Responsive 1.No Mandatory Certificates Available.
66	Ijaz enterprises	HFNC Model:not Mentioned Make:Guangdong China	Yes	No CeMDD, No USFDA, No MHLW	3+2 years	Yes	251	101 Units in 7-14 Days remaining in 15 to 25 days	Non Responsive 1.No Mandatory Certificates Available.

67	Ijaz enterprises	Bipap/Cpap Model:ST 125 Make:Hobi Topson Medical China	Yes	CEMDD,IS O 13485	3+2 years	Yes	500	500 in 7 to 15 Days	Responsive
68	BioCare Enterprises	ABG Machine with 15000 test and 10 Machines Model: i15 Make:Edan china	Yes	CEMDD,US FDA	3+2 years	Yes	10 machine with 15000	Ready Stock immediately after issue of PO	Responsive
69	BioCare Enterprises	ABG Machine Model: i15 Make:Edan china	Yes	CEMDD,US FDA	3+2 years	Yes	20	Ready Stock immediately after issue of PO	Responsive
70	Vertex Medical Pvt Ltd	Ventilator Model:Savina 300 Make:DragerWerk Germany	Yes	USFDA,CE MDD, ISO`s	3+2 years	Yes	186	60 Days	Responsive
71	Vertex Medical Pvt Ltd	Cardiac Monitor Model:Vista 120 Make:Dragerwerk China	Yes	USFDA,MH LW	3+2 years	Yes	906	60 Days	Responsive
72	Human Health Care	ABG Machine Model:Comblin Make:ESCHWEILER Germany	Yes	No CEMDD,No USFDA	3+2 years	Yes	10	Not Mentioned	Non Responsive 1.Certificate of Confirmity attached instead of CEMDD.

73	Human Health Care	ABG Machine with 15000 test Model: PT 1000 Make: easydiagnosis China	Yes	No CEMDD,N O USFDA	3+2 years	Yes	10	Not Mentioned	Non Responsive 1.Certificate of Confirmity attached instead of CEMDD.
74	Human Health Care	Cardiac monitor Model: K10 Make: Creative Medical China	Yes	CEMDD,US FDA	3+2 years	Yes	906	Not Mentioned	Responsive
75	Human Health Care	ECG Machine 3 Channel Model:1103B Make: Carewell China	Yes	CEMDD ,ISO	3+2 years	Yes	25	Not Mentioned	Non Responsive 1.The Qouted model do not have USFDA,MH LW. 2.Qouted Model is 1103B while The USFDA is to the model 1103G.
76	Human Health Care	Syringe Pump Model: Medifusion SP Make: Medicinos Gija UAB Europe	Yes	CEMDD,IS O	3+2 years	Yes	62	Not Mentioned	Responsive

77	Sudais Associates	Cardiac monitor Model:K10 Make:Creative China	Yes	CEMDD,US FDA	3+2 years	Yes	Nil	25 working days	Responsive
78	Sudais Associates	ECG Machine 3 Channel Model:IE-300 Make:Shenzhen Biocare China	Yes	CEMDD,US FDA	3+2 years	Yes	Nil	14 working days	Responsive
79	Sudais Associates	Patient trolley with portable Oxygen Cylinder Model:MER054-ET Make:Mermaid Industry China	Yes	ISO	3+2 years	Yes	Nil	28 working days	Non Responsive 1.mandatory Certificates not available.
80	Sudais Associates	Syringe Pump Model:IP 21-B Make:Biocare China	Yes	CEMDD	3+2 years	Yes	Nil	17 working days	Responsive
81	Sudais Associates	Bipap/Cpap Model:Bipap s9600 Make:Superstar China	Yes	CEMDD	3+2 years	Yes	Nil	14 working Days	Responsive
82	Platinum Corporation	Cardiac Monitor Model: Up-7000 Make:Creative China	Yes	CEMDD,US FDA	3+2 years	Yes	Nil	20 Days	Responsive
83	Indus Traders	Cardiac Monitor Model:critiview 360 Make:MDX-USA CHINA	Yes	No CEMDD,IS O	3+2 years	Yes	80	Remaining in 6 weeks	Non Responsive 1.Two out of Three Mandatory Certificates

									not Available.
84	Indus Traders	Cardiac Monitor Model: UPMED C500 Make:LE Medical CHINA	Yes	CEMDD	3+2 years	Yes	80	Remaining in 6 weeks	Non Responsive 1.Two out of Three Mandatory Certificates not Available.
85	Indus Traders	Cardiac Monitor Model: CS-12-Yonker Make:Carespace CHINA	Yes	CEMDD	3+2 years	Yes	80	Remaining in 6 weeks	Non Responsive 1.Two out of Three Mandatory Certificates not Available.
86	Indus Traders	Syringe Pump Model;Unifusion Sp50 Make:Medrena China	Yes	CEMDD	3+2 years	Yes	60	Remaining in 6 weeks	Responsive
87	Indus Traders	Bipap/Cpap Model: U25T Make:BMC China	Yes	CEMDD	3+2 years	Yes	50	Remaining in 6 weeks	Responsive
88	Indus Traders	Bipap/Cpap Model: ST30 Make: Ventmed China	Yes	CEMDD	3+2 years	Yes	150	Remaining in 6 weeks	Responsive

89	Meditec Instruments Co	ABG Machine Model:Not Mentioned Make:Medica USA	Yes	No USFDA, No CE, No MHLW	3+2 years	Yes	Null	Not Mentioned	Non Responsive 1.Mandatory Certificates two out of three not available
90	Alam Medix	Patient trolley with portable Oxygen Cylinder Model:CQ08 Make:MT-Medical China	Yes	CEMDD, IS O's	3+2 years	Yes	As per Requirement	within 2 hours	Responsive
91	Alam Medix	HFNC Model: HF6 Make:MT-Medical China	Yes	CEMDD, IS O's	3+2 years	Yes	Null	50 Days	Responsive
92	Alam Medix	ECG Machine 3 Channel Model: E30 G Make:Farum Poland	Yes	CEMDD	3+2 years	Yes	25	2 Hours	Non Responsive 1.Two out of Three Mandatory Certificates not Available.
93	Alam Medix	Portable X ray machine Model:MX-40 Make:Huestis USA Country of Origin:China	Yes	CEMDD, US FDA	3+2 years	Yes	Null	within 60 days	Responsive
94	ARNECA	Nothing Attached in the bid							Only company

									profile Attached
95	Universal Enterprises	Ventilator Model: Iternis advance Make: Heyer Germany	Yes	CEMDD	3+2 years	Yes	Null	90 days	Non Responsive 1. Two out of Three Mandatory Certificates not Available.
96	Universal Enterprises	Defibrillator Model: Cardio start Make: US DEFIB USA	Yes	USFDA, CE MDD	3+2 years	Yes	Null	90 days	Responsive
97	Universal Enterprises	Cardiac Monitor Model: Up-7000 Make: Creative China	Yes	USFDA, CE MDD	3+2 years	Yes	Null	90 days	Responsive
98	Universal Enterprises	ECG Machine 3 channel Model: 1103L Make: Carewell China	Yes	CEMDD	3+2 years	Yes	Null	90 days	Non Responsive 1. Two out of three Mandatory Certificates not Available
99	Universal Enterprises	Syringe Pump Model: SS 730 Make: Terumo Japan	Yes	MHLW, CE	3+2 years	Yes	Null	90 days	Responsive
100	Universal Enterprises	Flow Meter Model: FM20.XX Make: Inspital Turkey	Yes	CEMDD	3+2 years	Yes	Null	90 days	Responsive

101	Medical Equipment & System	Ventilator Model:EMV + 731 series Make: Zoll Medical USA	No	CEMDD	3+2 years	Yes	100	within Few Hours	Non Responsive 1.Hunadifier not qouted. 2.Qouted Modes are less than the required.
102	Medical Equipment & System	Defibrillator Model:M2 Make:zoll china	Yes	CEMDD,US FDA	3+2 years	Yes	21	within Few Hours	Responsive
103	Medical Equipment & System	Cardiac monitor Model:QUBE Compact Make:Spacelab USA	Yes	CEMDD,US FDA	3+2 years	Yes	Nil	6-8 weeks	Responsive
104	Medical Equipment & System	Cardiac Monitor Model:C-50 Make:Spacelab china	Yes	CEMDD	3+2 years	Yes	35	35 ready stock 871 in 6-8 weeks	Non Responsive 1.Two out of Three Mandatory Certificates not Available.
105	Medical Equipment & System	Syringe Pump Model:Infusia sp 7s Make:Fresinius Kabi China	Yes	No Mandatory Certificate	3+2 years	Yes	Nil	6-8 weeks	Non Responsive 1.mandatory Certificates not available. 2.Certificate of confirmity attached instead of CEMDD.

106	IBS Pharceutica 1	Semi Automated Chemistry analyzer Model:BiochemSA(BC-3000- C) Make:High Technology USA	Yes	No CEMDD,N O USFDA	3+2 years	Yes	Not Mentioned	Not Mentioned	Non Responsive 1.mandatory Certificates not available. 2.Certificate of conformity attached instead of CEMDD.
107	IBS Pharceutica 1	BIPAP/CPAP Model:Sleep one bi level ST Make:Kare medical Turkey	Yes	CEMDD	3+2 years	Yes	Not Mentioned	Not Mentioned	Responsive